

Visy Pulp and Paper Tumut CEMS - Correction Log

13/05/2020

Reporting Period: 1/04/2020 - 1/05/2020

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Hydrogen Chloride (HCl)		Period: 60 Minutes	Limit: 50.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
19/04/20 06:00	19/04/20 07:00	Instrument Error	Normal (Steady State)	Issues with the CEMS unit	53.14

Nitrogen Oxides (as NO2)		Period: 60 Minutes	Limit: 400.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
19/04/20 08:00	19/04/20 17:00	Instrument Error	Normal (Steady State)	Issues with the CEMS unit	545.07
19/04/20 17:00	19/04/20 19:00	Instrument Error	Normal (Steady State)	Issues with the CEMS unit	541.86
23/04/20 07:00	23/04/20 09:00	Instrument Error	Normal (Steady State)	Issues with the CEMS unit	667.86

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
12/04/20 06:06	12/04/20 06:24	Instrument Error	Normal (Steady State)	CEMS System Failure	90.62
12/04/20 20:06	12/04/20 20:12	Instrument Error	Normal (Steady State)	CEMS System Failure	20.98

Main Stack 2

Monitoring Location No: 22
Monitoring Type Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

TRS (as H2S)		Period: 60 Minutes	Limit: 3.60 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
12/04/20 02:00	12/04/20 03:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	3.86
19/04/20 02:00	19/04/20 03:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	4.35

Authorised By:

Uday Bhagwat
Pulp Mill Manager



15.5.20

Johan Stoltz
General Manager

